

IMPORTANT INFORMATION TO READ and RETURN

Installation Requirements for a Whitley A135 GMP Anaerobic Workstation

Thank you for choosing one of our products for your laboratory. To enable our engineers to perform an efficient, trouble-free installation please study, complete and email this form to us at **service@dwscientific.co.uk**. Should you have any questions, please do not hesitate to contact us, as we are here to help. When we have received the completed form, our Service Department will contact you to arrange a mutually convenient installation date.

The following information represents the ideal requirement.

Please contact us IMMEDIATELY if your intended location does not match this specification.

	Access Requ	irements							
	For access, the dimensions below should be taken into account when checking the size of doorways, lifts, stairs, etc								
	Space Requi	rements							
	The weight of the equipment is 295kg. If bench mounted, the bench allocated must be flat, level and of sufficient size to support the base fully.								
	External Dimensions*								
	Widt	h (mm)	Depth (mm)	Height (n	nm)				
	1:	500	1130	1080					
	* Please Note:								
	In addition to the dimensions stated above, allow for a localised protrusion of 120mm at rear to accommodate fans, condenser bottle, bracket, gas and electrical supplies. If necessary, this protrusion can be temporarily removed to fit through doors, etc.								
	If bench mounted, a further minimum clearance of 500mm is required above the unit and a minimum clearance of 500mm at the right hand side of the unit is required for user/service access.								
L	<u>'</u>								
	Gas Requirements The incoming gas supplies must be terminated near the right-hand side of the main chamber and fitted with leak-proof taps and pressure gauges.								
	The gas lines to which the equipment is attached are the responsibility of the user and should be constructed, tested and maintained to								
	the standards specified within the British Compressed Gasses Association (BCGA) Code of Practice CP4 (or international equivalent). Gas lines previously used for flammable gases must be purged prior to re-use.								
	Regulators should be fitted in accordance with the information contained in the table below and the various pressures strictly adhered to.								
	Gas Type	Connection	Details	Cylinder Regulator Required	Regulator Outlet Range	Flow Rate			
	Anaerobic Gas Mix		e fitting or connection vurethane tubing.	Two Stage	4 - 6 bar (60-90 psi)	60 litres per minute (dynamic)			
		Hydrogen/An							

DWS/SSP0005 Revision Level: 07

Two Stage

Notes: If the instrument is to be connected to a single cylinder of anaerobic gas mixture only, then this must be

4 - 6 bar (60-90 psi)

Minimum 150 litres

per minute (dynamic)

Stage – order Code A01745

for 8mm Polyurethane tubing

Nitrogen Regulator – Two Stage – order Code **A01748**

capable of delivering 150 litres/minute (sleeved ports).

Nitrogen

1/4" BSP male fitting or connection

Suitable Connection Types	Push in connection (Fittings of choice for DWS. Supplied with DWS spares kit). Push in connection (Customers' own preference. Not supplied by DWS).								
Mains Requirements									
Electricity Supply	Wall Socket	\neg							
240 Volts	1 x Three Pin, 13 Amp. Minimum 6A rating								
Other Considerations									
Although the workstation should be located in a well-ventilated area, avoid close proximity to air conditioning systems and draughts caused by windows and doors.									
Remember, if you do not have the required regulators, you can order them from Don Whitley Scientific: Hydrogen/Anaerobic Gas Mixture Regulator – Two Stage – order Code A01745 Nitrogen Regulator – Two Stage – order Code A01748									
Decontamination and Removal									
If an existing unit is being taken in part exchange or is being removed from the laboratory, it must be de-contaminated before DWS staff handle the unit. A certificate or signed letter confirming the unit has been decontaminated must be given to our engineer.									
There is a £400 fee for DWS to remove an existing unit from site. Please tick to accept this charge and an invoice will be provided.									
In the UK, delivery and installation are free of charge (unless otherwise agreed). If our engineers are unable to install the unit and a return journey is necessary, a charge may be made. Export customers, please refer to your local distributor. It is essential that this form is completed and returned, to avoid delay to your installation.									
THANK YOU FOR THINKING WHITLEY									

DWS/SSP0005 Revision Level: 07

Signature

Print Name

Title

Establishment